
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000608
9:16 09/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:41
STERILIZING 4:44
136.0 C 220kPa 4:44
136.1 C 220kPa 5:14
136.4 C 225kPa 5:44
136.3 C 221kPa 6:14
136.2 C 223kPa 6:44
136.4 C 223kPa 7:14
136.1 C 220kPa 7:44
136.3 C 222kPa 8:14
136.3 C 222kPa 8:14
VENTING 8:14
AIR DRYING 9:06
CYCLE COMPLETE 9:54

Digital Signature #
F619E674683151CA
