

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2165

Name: RITTER2  
Sterilizer ID: V2313228  
03/22/2023 09:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.2 F 0.0  
2:00 138.6 F 0.5  
4:00 150.7 F 3.0  
6:00 176.2 F 7.2  
8:00 200.9 F 12.5  
10:00 220.0 F 7.0  
12:00 233.1 F 12.2  
14:00 234.3 F 9.2  
16:00 237.9 F 9.7  
18:00 246.7 F 14.1  
20:00 259.2 F 20.9  
22:00 270.5 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.6 F 29.8  
2:00 272.1 F 29.2  
3:00 271.0 F 28.9  
4:00 272.9 F 30.2

Min 271.0 F 28.8  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.1 F 0.0  
5:00 208.7 F 0.0  
10:00 215.1 F 0.0  
15:00 217.6 F 0.0  
20:00 218.1 F 0.0  
25:00 220.1 F 0.0  
30:00 220.1 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 22:51  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:57  
CYCLE COMPLETE

03/22/2023 10:20 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_