

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000501  
7:00 22/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:56  
PRESSURIZING 3:57  
STERILIZING 5:32  
136.4 C 223kPa 5:32  
136.4 C 221kPa 6:02  
136.4 C 221kPa 6:32  
136.4 C 224kPa 7:02  
136.3 C 223kPa 7:32  
136.4 C 223kPa 8:02  
136.5 C 224kPa 8:32  
136.3 C 222kPa 9:02  
136.3 C 222kPa 9:02  
VENTING 9:02  
AIR DRYING 10:01  
CYCLE COMPLETE 12:29

-----  
Digital Signature #  
09648A523A5D62C5  
-----