
Midmark M11 - v1.0.0
Total Cycles: 1704

Name: RITTER 1
Sterilizer ID: V2178198
03/10/2022 07:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 75.8 F 0.0
4:00 82.7 F 0.1
6:00 120.4 F 1.0
8:00 149.1 F 2.4
10:00 169.2 F 4.4
12:00 189.7 F 7.7
14:00 208.8 F 11.3
16:00 221.1 F 4.3
18:00 229.2 F 7.6
20:00 238.2 F 11.7
22:00 233.8 F 7.9
24:00 246.0 F 13.5
26:00 243.3 F 11.8
28:00 255.4 F 18.0
30:00 265.8 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.6 F 28.9
2:00 271.3 F 28.1
3:00 272.4 F 29.1
4:00 272.5 F 28.7

Min 271.0 F 28.0
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 205.7 F 0.0
10:00 193.6 F 0.0
15:00 216.0 F 0.0
20:00 203.5 F 0.0
25:00 223.3 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:09:47
CYCLE COMPLETE

03/10/2022 09:24 AM
ACCEPTED: X
REJECTED:
Operator: U02