
Midmark M11 - v1.0.0
Total Cycles: 3184

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2023 07:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.0 F 0.0
2:00 76.5 F 0.0
4:00 83.7 F 0.0
6:00 108.7 F 0.1
8:00 144.0 F 1.9
10:00 173.3 F 4.2
12:00 209.4 F 7.6
14:00 211.2 F 11.6
16:00 223.9 F 4.5
18:00 229.1 F 8.9
20:00 239.3 F 13.2
22:00 239.4 F 10.2
24:00 236.0 F 8.5
26:00 248.3 F 14.4
28:00 259.7 F 20.7
30:00 270.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.8 F 28.3
3:00 271.8 F 28.7
4:00 272.5 F 28.9

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 156.6 F 0.0
10:00 199.2 F 0.0
15:00 198.7 F 0.0
20:00 204.5 F 0.0
25:00 220.3 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:56
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:00
CYCLE COMPLETE

05/02/2023 08:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____