
Midmark M11 - v1.0.0
Total Cycles: 3363

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2023 07:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.6 F 0.0
2:00 78.1 F 0.0
4:00 100.7 F 0.0
6:00 148.5 F 0.0
8:00 170.6 F 0.0
10:00 187.2 F 1.9
12:00 202.3 F 5.1
14:00 215.1 F 8.1
16:00 226.2 F 11.5
18:00 225.2 F 5.4
20:00 238.2 F 10.4
22:00 236.4 F 8.4
24:00 244.8 F 12.5
26:00 241.2 F 10.8
28:00 253.6 F 16.9
30:00 264.1 F 23.1
32:00 271.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.6 F 28.7
2:00 271.4 F 27.9
3:00 272.4 F 28.8
4:00 272.2 F 28.3

Min 271.2 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 190.9 F 0.0
10:00 204.3 F 0.0
15:00 222.5 F 0.0
20:00 232.3 F 0.0
25:00 229.2 F 0.0
30:00 227.9 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:10:09
CYCLE COMPLETE

06/09/2023 08:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____