
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001494
8:00 14/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:55
STERILIZING 3:56
136.0 C 220kPa 3:56
136.1 C 221kPa 4:26
136.2 C 222kPa 4:56
136.0 C 221kPa 5:26
136.4 C 224kPa 5:56
136.2 C 219kPa 6:26
136.4 C 220kPa 6:56
136.3 C 221kPa 7:26
136.3 C 221kPa 7:26
VENTING 7:26
AIR DRYING 8:16
CYCLE COMPLETE 8:20

Digital Signature #
A2FF8CBFF19F320F
