

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2056

Name: RITTER2  
Sterilizer ID: V2313228  
02/15/2023 01:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.3 F 0.0  
2:00 126.2 F 0.0  
4:00 155.1 F 0.4  
6:00 178.9 F 3.6  
8:00 202.7 F 7.7  
10:00 217.2 F 12.3  
12:00 224.4 F 7.5  
14:00 236.6 F 12.4  
16:00 236.1 F 10.0  
18:00 236.7 F 9.0  
20:00 248.4 F 15.0  
22:00 260.1 F 21.6  
24:00 270.9 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.6 F 29.6  
2:00 271.5 F 29.3  
3:00 272.6 F 29.6  
4:00 271.4 F 29.1

Min 271.4 F 29.0  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 202.3 F 0.0  
10:00 218.1 F 0.0  
15:00 223.7 F 0.0  
20:00 223.9 F 0.0  
25:00 225.7 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 24:44  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:02:51  
CYCLE COMPLETE

02/15/2023 02:22 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_