
Midmark M11 - v1.0.5
Total Cycles: 3618

Name: RITTER2
Sterilizer ID: V2313228
05/15/2024 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 70.2 F 0.0
4:00 73.7 F 0.0
6:00 105.0 F 0.3
8:00 141.2 F 1.2
10:00 166.5 F 3.6
12:00 185.8 F 6.5
14:00 209.3 F 9.7
16:00 221.2 F 13.3
18:00 216.3 F 7.1
20:00 231.1 F 11.1
22:00 229.4 F 7.4
24:00 241.3 F 12.4
26:00 239.5 F 10.6
28:00 251.2 F 16.4
30:00 261.6 F 22.3
32:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.4 F 29.6
2:00 271.3 F 28.8
3:00 272.4 F 29.7
4:00 271.7 F 29.0

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 227.6 F 0.0
10:00 188.3 F 0.0
15:00 185.1 F 0.0
20:00 182.7 F 0.0
25:00 181.7 F 0.0
30:00 211.1 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 32:43
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:11:01
CYCLE COMPLETE

05/15/2024 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____