
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001413
11:52 31/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:32
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 225kPa 4:08
136.4 C 224kPa 4:38
136.4 C 223kPa 5:08
136.0 C 220kPa 5:38
136.3 C 224kPa 6:08
136.4 C 224kPa 6:38
136.2 C 222kPa 7:08
136.2 C 222kPa 7:08
VENTING 7:08
AIR DRYING 7:56
CYCLE COMPLETE 22:56

Digital Signature #
B3194BB7496581D4
