

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005446  
9:50 01/05/2023

WRAPPED

135 C for 10 min.

CYCLE START	0:00
CONDITIONING	1:49
PRESSURIZING	3:47
STERILIZING	4:37
136.5 C 221kPa	4:37
136.3 C 222kPa	5:07
136.4 C 221kPa	5:37
136.3 C 223kPa	6:07
136.4 C 223kPa	6:37
136.3 C 222kPa	7:07
136.4 C 222kPa	7:37
136.4 C 224kPa	8:07
136.3 C 221kPa	8:37
136.4 C 217kPa	9:07
136.3 C 221kPa	9:37
136.4 C 224kPa	10:07
136.4 C 224kPa	10:37
136.4 C 225kPa	11:07
136.4 C 218kPa	11:37
136.2 C 224kPa	12:07
136.2 C 220kPa	12:37
136.2 C 217kPa	13:07
136.4 C 225kPa	13:37
136.4 C 220kPa	14:07
136.2 C 223kPa	14:37
136.2 C 223kPa	14:37
VENTING	14:37
AIR DRYING	15:25
CYCLE COMPLETE	24:06

-----

Digital Signature #

313789E9EA4F6519

-----