

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4674

Name: RITTER 1  
Sterilizer ID: V2178198  
06/25/2024 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.4 F 0.0  
2:00 96.1 F 0.6  
4:00 120.0 F 2.4  
6:00 149.4 F 5.0  
8:00 169.0 F 8.2  
10:00 190.9 F 12.0  
12:00 206.0 F 3.3  
14:00 216.6 F 6.5  
16:00 226.8 F 10.1  
18:00 231.8 F 9.5  
20:00 237.4 F 10.5  
22:00 235.9 F 8.4  
24:00 246.9 F 13.8  
26:00 257.7 F 19.6  
28:00 267.6 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.7 F 29.1  
2:00 271.5 F 28.5  
3:00 272.8 F 29.1  
4:00 271.6 F 28.3

Min 271.4 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 209.2 F 0.0  
10:00 220.0 F 0.0  
15:00 223.7 F 0.0  
20:00 218.1 F 0.0  
25:00 214.8 F 0.0  
30:00 212.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 29:27  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:07:56  
CYCLE COMPLETE

06/25/2024 12:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_