
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001548
11:45 11/01/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:04
PRESSURIZING 5:59
STERILIZING 7:44
136.0 C 220kPa 7:44
136.4 C 222kPa 8:14
136.5 C 222kPa 8:44
136.3 C 222kPa 9:14
136.4 C 220kPa 9:44
136.2 C 220kPa 10:14
136.4 C 221kPa 10:44
136.3 C 222kPa 11:14
VENTING 11:14
AIR DRYING 12:17
CYCLE COMPLETE 33:49

Digital Signature #
20FBDE6709D957B2
