
Midmark M11 - v1.0.0
Total Cycles: 1036

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2021 09:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.1 F 0.0
2:00 157.4 F 1.5
4:00 171.7 F 4.6
6:00 192.7 F 8.6
8:00 213.2 F 12.9
10:00 222.5 F 6.3
12:00 236.7 F 11.6
14:00 235.5 F 9.0
16:00 239.1 F 9.7
18:00 248.0 F 14.1
20:00 259.9 F 20.8
22:00 270.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 28.9
2:00 271.6 F 28.3
3:00 272.7 F 29.1
4:00 271.9 F 28.4

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 222.3 F 0.0
10:00 204.8 F 0.0
15:00 203.1 F 0.0
20:00 211.2 F 0.0
25:00 226.8 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

07/21/2021 10:29 AM
ACCEPTED: X
REJECTED:
Operator: U01