
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002212
15:33 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:42
STERILIZING 3:53
136.0 C 222kPa 3:53
136.2 C 224kPa 4:23
136.2 C 224kPa 4:53
136.4 C 226kPa 5:23
136.2 C 224kPa 5:53
136.4 C 226kPa 6:23
136.2 C 224kPa 6:53
136.4 C 227kPa 7:23
136.4 C 227kPa 7:23
VENTING 7:23
AIR DRYING 8:19
CYCLE COMPLETE 25:40

Digital Signature #
7437C17FE8CAD5C9
