
Midmark M11 - v1.0.0
Total Cycles: 5158

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2024 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.0 F 0.0
2:00 132.4 F 0.0
4:00 156.8 F 0.2
6:00 204.5 F 0.5
8:00 212.1 F 1.0
10:00 223.2 F 4.1
12:00 230.7 F 7.7
14:00 241.0 F 11.8
16:00 236.1 F 8.7
18:00 242.4 F 11.1
20:00 246.1 F 13.1
22:00 244.6 F 12.4
24:00 256.6 F 18.7
26:00 267.3 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.1 F 28.6
2:00 272.7 F 29.0
3:00 272.7 F 28.8
4:00 272.4 F 28.6

Min 271.6 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 232.3 F 0.0
10:00 229.0 F 0.0
15:00 222.7 F 0.0
20:00 223.8 F 0.0
25:00 220.9 F 0.0
30:00 212.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 27:40
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:06:16
CYCLE COMPLETE

12/04/2024 04:42 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____