
Midmark M11 - v1.0.5
Total Cycles: 988

Name: RITTER2
Sterilizer ID: V2313228
04/14/2022 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.2 F 0.0
2:00 98.9 F 0.0
4:00 109.0 F 0.0
6:00 150.6 F 0.4
8:00 185.7 F 1.6
10:00 204.8 F 5.4
12:00 216.2 F 9.7
14:00 230.0 F 9.3
16:00 234.9 F 10.4
18:00 235.4 F 8.5
20:00 242.3 F 11.4
22:00 246.5 F 13.9
24:00 259.0 F 20.7
26:00 270.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.3 F 29.7
2:00 271.5 F 28.8
3:00 272.5 F 29.5
4:00 271.4 F 28.8

Min 271.3 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 210.5 F 0.0
10:00 206.1 F 0.0
15:00 197.6 F 0.0
20:00 189.9 F 0.0
25:00 198.5 F 0.0
30:00 211.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 26:43
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:04:58
CYCLE COMPLETE

04/14/2022 11:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____