

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001409  
7:31 09/09/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:35  
PRESSURIZING 6:17  
STERILIZING 7:45  
136.0 C 220kPa 7:45  
136.4 C 216kPa 8:15  
136.4 C 221kPa 8:45  
136.3 C 220kPa 9:15  
136.4 C 222kPa 9:45  
136.0 C 219kPa 10:15  
136.0 C 219kPa 10:45  
136.4 C 222kPa 11:15  
VENTING 11:15  
AIR DRYING 12:18  
CYCLE COMPLETE 12:37  
-----

Digital Signature #  
11832C79A134F7FE  
-----