
Midmark M11 - v1.0.0
Total Cycles: 2819

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2023 11:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.4 F 0.0
2:00 191.2 F 1.7
4:00 199.0 F 6.5
6:00 218.4 F 11.9
8:00 228.6 F 7.7
10:00 241.4 F 13.2
12:00 239.3 F 10.7
14:00 237.0 F 9.0
16:00 251.5 F 16.0
18:00 263.4 F 22.9
20:00 271.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.4 F 28.7
2:00 271.3 F 28.0
3:00 272.5 F 29.2
4:00 272.2 F 28.6

Min 271.0 F 27.9
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 225.7 F 0.0
10:00 231.6 F 0.0
15:00 229.9 F 0.0
20:00 231.0 F 0.0
25:00 228.9 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:05
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

01/25/2023 12:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____