
Midmark M11 - v1.0.0
Total Cycles: 1193

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2021 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.1 F 0.0
2:00 199.1 F 3.0
4:00 198.3 F 8.1
6:00 217.6 F 10.9
8:00 228.9 F 10.0
10:00 232.7 F 7.5
12:00 243.9 F 13.4
14:00 243.3 F 11.9
16:00 256.5 F 18.7
18:00 268.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 29.3
2:00 272.4 F 29.0
3:00 272.0 F 29.0
4:00 272.2 F 29.3

Min 271.5 F 28.9
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.0 F 0.0
5:00 236.2 F 0.0
10:00 226.6 F 0.0
15:00 229.5 F 0.0
20:00 229.9 F 0.0
25:00 229.6 F 0.0
Dry Cycle Aborted

FILLING: 1:50
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:04
DRYING: 28:39
TOTAL CYCLE: 00:57:21

09/22/2021 10:24 AM
ACCEPTED: X
REJECTED:
Operator: U01