

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1416

Name: RITTER2  
Sterilizer ID: V2313228  
08/25/2022 12:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.6 F 0.0  
2:00 198.1 F 3.2  
4:00 213.2 F 8.2  
6:00 230.9 F 11.5  
8:00 234.0 F 10.4  
10:00 236.1 F 8.0  
12:00 240.8 F 10.9  
14:00 247.3 F 14.3  
16:00 260.3 F 21.5  
18:00 271.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.5 F 29.5  
2:00 271.5 F 29.1  
3:00 272.6 F 29.6  
4:00 271.7 F 28.9  
  
Min 271.5 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 228.2 F 0.0  
10:00 233.3 F 0.0  
15:00 224.4 F 0.0  
20:00 221.7 F 0.0  
25:00 223.8 F 0.0  
30:00 222.0 F 0.0  
  
Drying Complete

FILLING: 2:07  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:57:48  
CYCLE COMPLETE

08/25/2022 01:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_