
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001558
12:26 26/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:11
STERILIZING 4:02
136.0 C 220kPa 4:02
136.3 C 219kPa 4:32
136.0 C 222kPa 5:02
136.2 C 222kPa 5:32
136.2 C 220kPa 6:02
136.3 C 224kPa 6:32
136.3 C 224kPa 7:02
136.3 C 223kPa 7:32
136.3 C 223kPa 7:32
VENTING 7:32
AIR DRYING 8:25
CYCLE COMPLETE 8:35

Digital Signature #
ADF495D91E22DAE3
