
Midmark M11 - v1.0.0
Total Cycles: 2828

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2023 04:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.1 F 0.0
2:00 202.3 F 5.4
4:00 211.5 F 11.1
6:00 222.9 F 6.4
8:00 235.2 F 12.1
10:00 236.5 F 9.7
12:00 237.2 F 8.6
14:00 250.3 F 15.3
16:00 262.5 F 22.3
18:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 29.0
2:00 271.5 F 28.2
3:00 272.5 F 29.1
4:00 271.9 F 28.5

Min 271.4 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 230.7 F 0.0
10:00 237.9 F 0.0
15:00 237.9 F 0.0
20:00 229.1 F 0.0
25:00 225.4 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 18:13
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

01/30/2023 05:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____