
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001763
16:07 01/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:46
PRESSURIZING 1:53
STERILIZING 2:57
136.4 C 226kPa 2:57
136.4 C 226kPa 3:27
136.4 C 226kPa 3:57
136.4 C 226kPa 4:27
136.2 C 225kPa 4:57
136.2 C 225kPa 5:27
136.4 C 225kPa 5:57
136.3 C 225kPa 6:27
136.3 C 225kPa 6:27
VENTING 6:27
AIR DRYING 7:17
CYCLE COMPLETE 27:20

Digital Signature #
E1231416206F71AF
