
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001179
11:38 21/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:05
PRESSURIZING 4:06
STERILIZING 5:42
136.4 C 224kPa 5:42
136.2 C 223kPa 6:12
136.2 C 222kPa 6:42
135.9 C 221kPa 7:12
136.3 C 224kPa 7:42
135.9 C 222kPa 8:12
136.4 C 225kPa 8:42
136.5 C 221kPa 9:12
136.5 C 221kPa 9:12
VENTING 9:12
AIR DRYING 10:09
CYCLE COMPLETE 10:19

Digital Signature #
8600AA5B503B97C6
