
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004507
10:42 06/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:14
PRESSURIZING 1:19
STERILIZING 2:18
136.4 C 220kPa 2:18
136.4 C 224kPa 2:48
136.4 C 224kPa 3:18
136.4 C 223kPa 3:48
136.4 C 224kPa 4:18
136.3 C 224kPa 4:48
136.4 C 222kPa 5:18
136.3 C 221kPa 5:48
136.3 C 221kPa 5:48
VENTING 5:48
AIR DRYING 6:35
CYCLE COMPLETE 6:38

Digital Signature #
C9CAD9D6C6870643
