
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 002536
16:27 26/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:45
PRESSURIZING 0:50
STERILIZING 1:58
136.0 C 220kPa 1:58
136.4 C 221kPa 2:28
136.4 C 223kPa 2:58
136.4 C 223kPa 3:28
136.3 C 220kPa 3:58
136.4 C 223kPa 4:28
136.4 C 224kPa 4:58
136.2 C 223kPa 5:28
136.2 C 223kPa 5:28
VENTING 5:28
AIR DRYING 6:22
CYCLE COMPLETE 6:25

Digital Signature #
F7C843B02D63CF23
