

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001560  
11:28 06/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:49  
PRESSURIZING 3:50  
STERILIZING 4:58  
136.5 C 224kPa 4:58  
136.2 C 223kPa 5:28  
136.3 C 224kPa 5:58  
136.3 C 222kPa 6:28  
136.3 C 223kPa 6:58  
136.4 C 224kPa 7:28  
136.2 C 222kPa 7:58  
136.4 C 224kPa 8:28  
136.4 C 224kPa 8:28  
VENTING 8:28  
AIR DRYING 9:20  
CYCLE COMPLETE 9:32

-----  
Digital Signature #  
334754DE8DB59C05  
-----