
Midmark M11 - v1.0.5
Total Cycles: 4112

Name: RITTER2
Sterilizer ID: V2313228
11/13/2024 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.3 F 0.0
2:00 139.1 F 0.0
4:00 146.1 F 0.3
6:00 178.0 F 3.6
8:00 193.7 F 7.6
10:00 210.4 F 11.8
12:00 226.0 F 6.3
14:00 232.8 F 10.9
16:00 233.4 F 7.5
18:00 243.5 F 13.6
20:00 244.4 F 12.5
22:00 255.7 F 18.8
24:00 266.4 F 25.1
26:00 272.6 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.2 F 29.1
2:00 272.7 F 29.3
3:00 272.4 F 29.5
4:00 272.3 F 29.7

Min 271.7 F 29.1
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 235.6 F 0.0
10:00 190.9 F 0.0
15:00 187.6 F 0.0
20:00 184.4 F 0.0
25:00 184.0 F 0.0
30:00 183.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:23
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:56
CYCLE COMPLETE

11/13/2024 11:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____