
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002109
13:19 02/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 3:52
136.4 C 226kPa 3:52
136.4 C 226kPa 4:22
136.4 C 225kPa 4:52
136.2 C 224kPa 5:22
136.2 C 223kPa 5:52
136.4 C 226kPa 6:22
136.4 C 225kPa 6:52
136.4 C 226kPa 7:22
136.4 C 226kPa 7:22
VENTING 7:22
AIR DRYING 8:13
CYCLE COMPLETE 38:13

Digital Signature #
B42041785D9FD25A
