
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001889
8:19 28/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:42
STERILIZING 4:45
136.4 C 221kPa 4:45
136.3 C 222kPa 5:15
136.4 C 223kPa 5:45
136.2 C 220kPa 6:15
136.4 C 222kPa 6:45
136.1 C 222kPa 7:15
136.2 C 222kPa 7:45
136.4 C 223kPa 8:15
136.4 C 223kPa 8:15
VENTING 8:15
AIR DRYING 9:10
CYCLE COMPLETE 9:12

Digital Signature #
76FAAD72F78E8C4E
