

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2935

Name: RITTER2  
Sterilizer ID: V2313228  
09/27/2023 08:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 80.2 F 0.0  
4:00 121.6 F 0.0  
6:00 114.2 F 0.0  
8:00 175.6 F 0.0  
10:00 181.5 F 0.8  
12:00 207.3 F 4.4  
14:00 221.8 F 7.8  
16:00 232.9 F 11.5  
18:00 227.9 F 6.2  
20:00 239.1 F 10.9  
22:00 234.2 F 7.7  
24:00 246.0 F 12.9  
26:00 244.6 F 12.4  
28:00 256.2 F 18.6  
30:00 266.7 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 28.8  
2:00 272.0 F 28.9  
3:00 272.2 F 28.5  
4:00 272.3 F 29.2

Min 271.4 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
5:00 226.4 F 0.0  
10:00 228.1 F 0.0  
15:00 221.9 F 0.0  
20:00 229.5 F 0.0  
25:00 225.6 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:35  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:09:36  
CYCLE COMPLETE

09/27/2023 09:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_