
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008005
7:56 30/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:52
STERILIZING 3:50
136.4 C 221kPa 3:50
135.9 C 220kPa 4:20
135.7 C 218kPa 4:50
136.2 C 221kPa 5:20
136.0 C 218kPa 5:50
136.4 C 222kPa 6:20
136.4 C 221kPa 6:50
136.2 C 222kPa 7:20
136.2 C 222kPa 7:20
VENTING 7:20
AIR DRYING 8:16
CYCLE COMPLETE 38:16

Digital Signature #
6AAC016E68349C4A
