
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000995
11:08 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:16
STERILIZING 3:11
136.0 C 227kPa 3:11
136.3 C 233kPa 3:41
136.3 C 232kPa 4:11
136.1 C 231kPa 4:41
136.3 C 231kPa 5:11
136.3 C 233kPa 5:41
136.4 C 233kPa 6:11
136.4 C 233kPa 6:41
136.4 C 233kPa 6:41
VENTING 6:41
AIR DRYING 7:34
CYCLE COMPLETE 7:42

Digital Signature #
74750566B86A5DD6
