

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4097

Name: RITTER 1  
Sterilizer ID: V2178198  
12/08/2023 11:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.5 F 0.0  
2:00 155.7 F 1.3  
4:00 154.4 F 4.2  
6:00 171.7 F 7.4  
8:00 192.3 F 11.2  
10:00 215.4 F 3.5  
12:00 217.7 F 7.3  
14:00 228.6 F 11.3  
16:00 228.9 F 6.8  
18:00 240.7 F 12.2  
20:00 239.1 F 10.2  
22:00 251.5 F 16.1  
24:00 262.2 F 22.1  
26:00 271.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.4 F 29.1  
2:00 271.7 F 28.6  
3:00 271.4 F 28.6  
4:00 272.6 F 29.1

Min 271.1 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 224.2 F 0.0  
10:00 230.8 F 0.0  
15:00 231.2 F 0.0  
20:00 228.1 F 0.0  
25:00 230.3 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 26:37  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:05:28  
CYCLE COMPLETE

12/08/2023 12:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_