
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001593
14:24 15/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:34
STERILIZING 3:34
136.0 C 220kPa 3:34
136.2 C 223kPa 4:04
136.3 C 225kPa 4:34
136.4 C 226kPa 5:04
136.4 C 225kPa 5:34
136.4 C 225kPa 6:04
136.4 C 225kPa 6:34
136.4 C 225kPa 7:04
136.4 C 225kPa 7:04
VENTING 7:04
AIR DRYING 7:55
CYCLE COMPLETE 8:00

Digital Signature #
BAFB07F1C9CF6FEE
