

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1430

Name: RITTER 1  
Sterilizer ID: V2178198  
12/08/2021 07:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.6 F 0.0  
2:00 72.4 F 0.0  
4:00 76.9 F 0.0  
6:00 94.3 F 0.0  
8:00 126.2 F 0.2  
10:00 170.5 F 0.3  
12:00 192.5 F 0.5  
14:00 208.6 F 3.6  
16:00 221.5 F 6.3  
18:00 229.2 F 9.7  
20:00 241.0 F 11.7  
22:00 238.0 F 10.3  
24:00 238.7 F 8.3  
26:00 244.9 F 11.8  
28:00 245.9 F 12.9  
30:00 258.0 F 19.4  
32:00 268.6 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.0 F 28.2  
2:00 272.5 F 28.7  
3:00 271.8 F 28.4  
4:00 271.8 F 28.1  
  
Min 271.4 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 227.3 F 0.0  
10:00 237.4 F 0.0  
15:00 220.6 F 0.0  
20:00 211.0 F 0.0  
25:00 206.1 F 0.0  
30:00 205.2 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 33:08  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:11:08  
CYCLE COMPLETE

12/08/2021 10:10 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02