
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000930
10:56 08/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:11
STERILIZING 4:16
136.0 C 220kPa 4:16
136.4 C 224kPa 4:46
136.4 C 223kPa 5:16
136.2 C 223kPa 5:46
136.1 C 217kPa 6:16
136.4 C 224kPa 6:46
136.2 C 224kPa 7:16
136.2 C 223kPa 7:46
136.2 C 223kPa 7:46
VENTING 7:46
AIR DRYING 8:40
CYCLE COMPLETE 8:49

Digital Signature #
4EBE22A6AFCBBEEF
