

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
2.0uS / 1.2ppm  
CYCLE NUMBER 000031  
13:14 10/08/2020  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 4:05  
STERILIZING 5:10  
136.0 C 223kPa 5:10  
136.2 C 226kPa 5:40  
136.0 C 225kPa 6:10  
136.4 C 228kPa 6:40  
136.4 C 227kPa 7:10  
136.4 C 227kPa 7:40  
136.3 C 227kPa 8:10  
136.4 C 225kPa 8:40  
VENTING 8:40  
AIR DRYING 9:34  
CYCLE COMPLETE 39:34  
-----

Digital Signature #  
E7C1A3E4CDE0255C  
-----