

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4924

Name: RITTER 1  
Sterilizer ID: V2178198  
09/20/2024 08:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.7 F 0.0  
2:00 70.7 F 0.0  
4:00 85.7 F 0.0  
6:00 115.5 F 0.0  
8:00 148.4 F 0.2  
10:00 177.0 F 0.3  
12:00 197.7 F 1.1  
14:00 217.6 F 4.5  
16:00 221.9 F 7.7  
18:00 231.4 F 11.2  
20:00 229.8 F 6.4  
22:00 241.4 F 11.7  
24:00 237.6 F 9.2  
26:00 239.0 F 9.3  
28:00 248.6 F 14.2  
30:00 260.2 F 20.7  
32:00 270.4 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 27.9  
1:00 271.6 F 28.0  
2:00 272.5 F 28.7  
3:00 272.8 F 28.7  
4:00 272.2 F 28.3  
  
Min 271.5 F 27.8  
Max 272.9 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.4 F 0.0  
5:00 207.4 F 0.0  
10:00 226.8 F 0.0  
15:00 206.1 F 0.0  
20:00 193.7 F 0.0  
25:00 199.8 F 0.0  
30:00 204.2 F 0.0  
  
Drying Complete

FILLING: 0:42  
HEATING: 32:50  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:10:58  
CYCLE COMPLETE

09/20/2024 09:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_