
Midmark M11 - v1.0.5
Total Cycles: 4091

Name: RITTER2
Sterilizer ID: V2313228
11/06/2024 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.2 F 0.0
4:00 78.1 F 0.0
6:00 100.7 F 0.0
8:00 148.9 F 0.1
10:00 181.4 F 0.4
12:00 201.0 F 2.4
14:00 211.5 F 5.5
16:00 219.9 F 8.8
18:00 227.8 F 12.5
20:00 230.6 F 8.2
22:00 241.5 F 13.0
24:00 239.9 F 11.0
26:00 237.3 F 9.5
28:00 250.9 F 16.0
30:00 261.9 F 22.4
32:00 272.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.7 F 29.4
2:00 271.5 F 28.7
3:00 272.7 F 29.4
4:00 271.7 F 28.6

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 208.2 F 0.0
10:00 192.1 F 0.0
15:00 188.7 F 0.0
20:00 187.7 F 0.0
25:00 192.9 F 0.0
30:00 198.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 32:32
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:47
CYCLE COMPLETE

11/06/2024 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____