
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001433
10:35 09/15/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:44
PRESSURIZING 6:28
STERILIZING 7:55
136.0 C 220kPa 7:55
135.9 C 217kPa 8:25
136.3 C 220kPa 8:55
136.0 C 218kPa 9:25
136.5 C 221kPa 9:55
136.4 C 220kPa 10:25
136.0 C 217kPa 10:55
136.2 C 221kPa 11:25
VENTING 11:25
AIR DRYING 12:27
CYCLE COMPLETE 42:27

Digital Signature #
11AA042FA511C8AC
