

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3414

Name: RITTER 1  
Sterilizer ID: V2178198  
06/20/2023 12:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
2:00 199.7 F 3.9  
4:00 214.8 F 8.7  
6:00 229.1 F 8.6  
8:00 234.1 F 10.3  
10:00 235.7 F 7.7  
12:00 243.4 F 11.8  
14:00 245.4 F 13.0  
16:00 258.3 F 19.9  
18:00 269.6 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.3  
1:00 272.7 F 29.1  
2:00 271.6 F 28.2  
3:00 272.6 F 29.0  
4:00 271.7 F 28.2  
  
Min 271.5 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 218.6 F 0.0  
10:00 232.0 F 0.0  
15:00 231.3 F 0.0  
20:00 224.3 F 0.0  
25:00 218.9 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 2:06  
HEATING: 18:52  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:06  
CYCLE COMPLETE

06/20/2023 01:11 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_