

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2272

Name: RITTER 1  
Sterilizer ID: V2178198  
08/31/2022 11:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.7 F 0.0  
2:00 129.5 F 0.1  
4:00 149.8 F 0.5  
6:00 179.5 F 2.1  
8:00 201.7 F 6.3  
10:00 219.4 F 10.8  
12:00 222.8 F 6.0  
14:00 234.0 F 10.6  
16:00 232.8 F 7.7  
18:00 244.7 F 13.3  
20:00 242.9 F 11.8  
22:00 255.6 F 18.2  
24:00 266.2 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 29.1  
2:00 271.7 F 28.3  
3:00 271.9 F 28.7  
4:00 272.6 F 28.9

Min 271.4 F 28.1  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.2 F 0.0  
5:00 212.4 F 0.0  
10:00 230.2 F 0.0  
15:00 231.6 F 0.0  
20:00 228.5 F 0.0  
25:00 232.1 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 25:46  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:03:49  
CYCLE COMPLETE

08/31/2022 12:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_