
Midmark M11 - v1.0.5
Total Cycles: 1855

Name: RITTER2
Sterilizer ID: V2313228
12/19/2022 12:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.7 F 0.0
2:00 200.6 F 4.7
4:00 206.4 F 10.5
6:00 221.6 F 6.7
8:00 236.5 F 13.2
10:00 239.7 F 11.7
12:00 240.5 F 11.3
14:00 255.1 F 18.8
16:00 267.7 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.2 F 29.5
2:00 271.8 F 29.7
3:00 272.2 F 29.5
4:00 271.7 F 29.6

Min 271.4 F 29.1
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 225.2 F 0.0
10:00 241.4 F 0.0
15:00 237.9 F 0.0
20:00 230.6 F 0.0
25:00 225.3 F 0.0
30:00 224.5 F 0.0

Drying Complete

FILLING: 2:30
HEATING: 17:13
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:56:54
CYCLE COMPLETE

12/19/2022 01:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____