

-----  
Midmark M11 - v1.0.5  
Total Cycles: 46

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2021 01:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.4 F 0.0  
2:00 126.8 F 1.6  
4:00 143.5 F 4.0  
6:00 170.9 F 7.4  
8:00 193.8 F 11.6  
10:00 208.9 F 4.5  
12:00 219.7 F 8.4  
14:00 231.4 F 12.7  
16:00 233.6 F 9.6  
18:00 238.7 F 10.5  
20:00 245.8 F 13.7  
22:00 256.6 F 19.5  
24:00 266.4 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.4 F 29.9  
2:00 271.4 F 29.2  
3:00 272.1 F 30.0  
4:00 271.7 F 29.4

Min 271.2 F 29.2  
Max 272.4 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
5:00 230.9 F 0.0  
10:00 226.2 F 0.0  
15:00 220.5 F 0.0  
20:00 190.1 F 0.0  
25:00 186.5 F 0.0  
30:00 182.5 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 25:45  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:04:22  
CYCLE COMPLETE

05/26/2021 02:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_