
Midmark M11 - v1.0.5
Total Cycles: 3946

Name: RITTER2
Sterilizer ID: V2313228
09/13/2024 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 178.2 F 2.8
4:00 186.1 F 7.0
6:00 204.4 F 11.6
8:00 218.6 F 5.6
10:00 230.5 F 10.2
12:00 233.7 F 8.4
14:00 243.5 F 13.1
16:00 243.3 F 12.2
18:00 255.7 F 18.6
20:00 266.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.7 F 29.3
2:00 272.1 F 29.4
3:00 272.4 F 29.1
4:00 272.2 F 29.5

Min 271.7 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 225.5 F 0.0
10:00 230.5 F 0.0
15:00 227.1 F 0.0
20:00 222.6 F 0.0
25:00 220.5 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

09/13/2024 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____