

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4012

Name: RITTER2  
Sterilizer ID: V2313228  
10/09/2024 08:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 70.4 F 0.0  
4:00 72.3 F 0.0  
6:00 95.3 F 0.1  
8:00 139.7 F 0.3  
10:00 162.2 F 0.7  
12:00 185.4 F 3.2  
14:00 206.0 F 6.0  
16:00 212.6 F 9.4  
18:00 220.8 F 13.0  
20:00 225.7 F 7.8  
22:00 235.7 F 12.1  
24:00 234.1 F 9.0  
26:00 242.3 F 11.5  
28:00 243.8 F 12.6  
30:00 255.6 F 18.8  
32:00 266.2 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.7 F 29.7  
2:00 271.6 F 28.9  
3:00 272.1 F 29.8  
4:00 272.4 F 29.4  
  
Min 271.3 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.3 F 0.0  
5:00 201.5 F 0.0  
10:00 189.0 F 0.0  
15:00 180.0 F 0.0  
20:00 178.6 F 0.0  
25:00 179.0 F 0.0  
30:00 179.8 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 33:43  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:12:12  
CYCLE COMPLETE

10/09/2024 10:05 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_