

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3661

Name: RITTER 1  
Sterilizer ID: V2178198  
08/16/2023 09:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.7 F 0.0  
2:00 171.7 F 1.8  
4:00 176.5 F 4.8  
6:00 208.8 F 8.7  
8:00 210.2 F 13.5  
10:00 223.4 F 7.1  
12:00 236.3 F 11.8  
14:00 235.1 F 8.5  
16:00 243.5 F 11.6  
18:00 244.5 F 12.5  
20:00 255.6 F 18.5  
22:00 266.3 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.6 F 29.2  
2:00 271.4 F 28.3  
3:00 272.3 F 29.3  
4:00 272.0 F 28.8

Min 271.1 F 28.3  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.3 F 0.0  
5:00 221.9 F 0.0  
10:00 223.1 F 0.0  
15:00 217.2 F 0.0  
20:00 197.7 F 0.0  
25:00 238.4 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 23:47  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:02:45  
CYCLE COMPLETE

08/16/2023 10:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_