
Midmark M11 - v1.0.5
Total Cycles: 3625

Name: RITTER2
Sterilizer ID: V2313228
05/16/2024 09:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.7 F 0.0
2:00 160.3 F 1.9
4:00 166.8 F 5.3
6:00 187.9 F 9.3
8:00 205.2 F 10.9
10:00 218.3 F 8.0
12:00 231.7 F 12.8
14:00 233.6 F 9.7
16:00 237.3 F 9.5
18:00 245.8 F 13.9
20:00 256.8 F 20.0
22:00 267.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.5 F 30.1
2:00 272.0 F 29.5
3:00 271.3 F 29.6
4:00 272.6 F 30.1

Min 271.1 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 224.4 F 0.0
10:00 227.0 F 0.0
15:00 231.0 F 0.0
20:00 230.9 F 0.0
25:00 229.4 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 23:35
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:23
CYCLE COMPLETE

05/16/2024 10:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____